

## Hygroscopic Tempilstiks®

The following temperature indicating sticks are subject to absorbing moisture over time. Once the stick has absorbed moisture it may become difficult to advance from its holder. Additionally, sticks that have absorbed moisture may result in compromised strength and temperature measurement accuracy.

### Tempilstik Temperature

### LA-CO Industries Part Number

700° F / 371° C	28049
750° F / 399° C	28050
800° F / 427° C	28051
850° F / 460° C	28052
900° F / 482° C	28053
950° F / 510° C	28055
1000° F / 538° C	28056
1022° F / 550° C	28057
1050° F / 566° C	28058
1100° F / 593° C	28059
1150° F / 621° C	28060
1200° F / 649° C	28061
1300° F / 704° C	28063
1400° F / 760° C	28064
1450° F / 788° C	28065
1700° F / 927° C	28070
500° C / 932° F	28349
560° C / 1040° F	28350
600° C / 1112° F	28351
625° C / 1157° F	28352
700° C / 1292° F	28355

The above temperature sticks are specially packed in plastic sleeves with desiccant. The Tempilstik products should be stored within this packaging prior to use. These specific products should be stored in areas with a low humidity environment. Ideally a temperature / humidity controlled room should be used with temperatures not exceeding 86° F (30° C) and humidity less than 60%. Care should be taken when opening the plastic sleeve in order to retain it and the desiccant for reuse when storing the Tempilstik within the controlled environment.